



SEA-BIRD ELECTRONICS, INC.
1808 - 136th Place Northeast, Bellevue, Washington 98005 USA
Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

ENTERED
FEB 04 2005

Pressure Test Certificate

Customer University of Miami

Job Number 37906

Date 12/28/2004

Technician BLT

Serial Number 041847

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 30 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.

