

ENTERED  
FEB 04 2005

**SBE** SEA-BIRD ELECTRONICS, INC.  
1808 - 136th Place Northeast, Bellevue, Washington 98005 USA  
Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer University of Miami  
Job Number 37906  
Date 12/15/2004  
Technician BLT

Serial Number 032652

Low Pressure (PSI) 50 PSI  
Time (Minutes) 15 Minutes  
High Pressure (PSI) 10000 PSI  
Time (Minutes) 30 Minutes

Pass   
Fail

Comments  
Replaced the main piston "O"-Rings.

