



*Sensing Technology*

## WeatherStation® Instrument Calibration Certificate

**Model:** PB200  
**ID:** 2265416  
**Calibration Date:** 6/2/2011  
**Manufacturer :** Airmar Technology Corporation, Milford NH USA  
**Test Procedure:** 96-154-02

The WeatherStation® Instrument passes testing for IPX6 heavy seas water standards (excepting units equipped with RH sensors, which meet IPX4) and the IEC 60945 standards for exposed units<sup>1</sup>. All units are calibrated in an onsite wind tunnel<sup>2</sup>.

### Unit Test Results:

Test Performed	Logged Value	Test Requirement
0 Knot Wind	0.9	Less than 1 Knot Max
50 Knot Cal <sub>4</sub>	PASS	Within 2 Knots RMS
Humidity	N/A	Within ±5 % RH
Temperature <sub>5</sub>	22.3	Within ±1.1 °C
Pressure	997.7	Within ±2 mBar
Rate Gyro – Dynamic <sub>6</sub>	N/A	Under 4 deg/sec average error
Rate Gyro – Static	N/A	Within ±40 Deg/min
Pitch	0.0	Within 1°
Roll	0.2	Within 1°

### Ambient (Local) Test Conditions<sub>3</sub>:

**Humidity:** 39.5 %RH  
**Temperature:** 22.7 Deg C  
**Pressure:** 997.9 mBars

### Equipment List

- **Wind Tunnel Calibration**
  - United Sensor Corp. Pitot Tube – Calibrated to NIST traceable pitot tube
  - MKS Instruments Pressure Sensor
  - BK Precision® 5491A Multimeter
- **Ambient Temperature Readings (Temperature, humidity and pressure)**
  - Vaisala® PTU200
- **Rate Gyro Testing**
  - SEI A2 Absolute encoder
  - US Digital R164 Motor Controller

- 1 – After multiple heat cycles above 65° C (149°F) wind anemometer performance may require recalibration to remain within specifications at wind speeds above 50 knots (92.6 km/h) .
- 2 – The on site wind tunnel is calibrated with a pitot tube, which in turn was calibrated with a NIST traceable pitot tube.
- 3 – Ambient conditions measured with a Vaisala PTU300. Temperature, humidity and pressure readings compared to Vaisala PTU300. The temperature, pressure and humidity readings from the Vaisala® Instruments were calibrated by Vaisala® against instruments calibrated to NIST traceable instruments.
- 4 – Units are calibrated at 50 knots with a 75 point correction curve.
- 5 – Unit temperature and humidity recorded with 4+ knots of wind present.
- 6 – Test conducted at a rate of turn of 70 degrees per second.



*Sensing Technology*

## WeatherStation® Instrument Calibration Certificate

**Model:** PB200  
**ID:** 2250051  
**Calibration Date:** 5/14/2011  
**Manufacturer :** Airmar Technology Corporation, Milford NH USA  
**Test Procedure:** 96-154-02

The WeatherStation® Instrument passes testing for IPX6 heavy seas water standards (excepting units equipped with RH sensors, which meet IPX4) and the IEC 60945 standards for exposed units<sup>1</sup>. All units are calibrated in an onsite wind tunnel<sup>2</sup>.

### Unit Test Results:

Test Performed	Logged Value	Test Requirement
0 Knot Wind	0.1	Less than 1 Knot Max
50 Knot Cal <sub>4</sub>	PASS	Within 2 Knots RMS
Humidity	N/A	Within ±5 % RH
Temperature <sub>5</sub>	22.8	Within ±1.1 °C
Pressure	1001.6	Within ±2 mBar
Rate Gyro – Dynamic <sub>6</sub>	N/A	Under 4 deg/sec average error
Rate Gyro – Static	N/A	Within ±40 Deg/min
Pitch	0.0	Within 1°
Roll	0.0	Within 1°

### Ambient (Local) Test Conditions<sup>3</sup>:

**Humidity:** 31.7 %RH  
**Temperature:** 23.8 Deg C  
**Pressure:** 1001.4 mBars

### Equipment List

- **Wind Tunnel Calibration**
  - United Sensor Corp. Pitot Tube – Calibrated to NIST traceable pitot tube
  - MKS Instruments Pressure Sensor
  - BK Precision® 5491A Multimeter
- **Ambient Temperature Readings (Temperature, humidity and pressure)**
  - Vaisala® PTU200
- **Rate Gyro Testing**
  - SEI A2 Absolute encoder
  - US Digital R164 Motor Controller

- 1 – After multiple heat cycles above 65° C (149°F) wind anemometer performance may require recalibration to remain within specifications at wind speeds above 50 knots (92.6 km/h) .
- 2 – The on site wind tunnel is calibrated with a pitot tube, which in turn was calibrated with a NIST traceable pitot tube.
- 3 – Ambient conditions measured with a Vaisala PTU300. Temperature, humidity and pressure readings compared to Vaisala PTU300. The temperature, pressure and humidity readings from the Vaisala® instruments were calibrated by Vaisala® against instruments calibrated to NIST traceable instruments.
- 4 – Units are calibrated at 50 knots with a 75 point correction curve.
- 5 – Unit temperature and humidity recorded with 4+ knots of wind present.
- 6 – Test conducted at a rate of turn of 70 degrees per second.

## WeatherStation® Instrument Calibration Certificate

**Model:** PB200  
**ID:** 2265420  
**Calibration Date:** 6/2/2011  
**Manufacturer :** Airmar Technology Corporation, Milford NH USA  
**Test Procedure:** 96-154-02

The WeatherStation® Instrument passes testing for IPX6 heavy seas water standards (excepting units equipped with RH sensors, which meet IPX4) and the IEC 60945 standards for exposed units<sub>1</sub>. All units are calibrated in an onsite wind tunnel<sub>2</sub>.

### Unit Test Results:

Test Performed	Logged Value	Test Requirement
0 Knot Wind	0.5	Less than 1 Knot Max
50 Knot Cal <sub>4</sub>	PASS	Within 2 Knots RMS
Humidity	N/A	Within ±5 % RH
Temperature <sub>5</sub>	22.7	Within ±1.1 °C
Pressure	998.1	Within ±2 mBar
Rate Gyro – Dynamic <sub>6</sub>	N/A	Under 4 deg/sec average error
Rate Gyro – Static	N/A	Within ±40 Deg/min
Pitch	0.0	Within 1°
Roll	0.2	Within 1°

### Ambient (Local) Test Conditions<sub>3</sub>:

**Humidity:** 39.4 %RH  
**Temperature:** 23.4 Deg C  
**Pressure:** 998.1 mBars

### Equipment List

- Wind Tunnel Calibration
  - United Sensor Corp. Pitot Tube – Calibrated to NIST traceable pitot tube
  - MKS Instruments Pressure Sensor
  - BK Precision® 5491A Multimeter
- Ambient Temperature Readings (Temperature, humidity and pressure)
  - Vaisala® PTU200
- Rate Gyro Testing
  - SEI A2 Absolute encoder
  - US Digital R164 Motor Controller

- 1 – After multiple heat cycles above 65° C (149°F) wind anemometer performance may require recalibration to remain within specifications at wind speeds above 50 knots (92.6 km/h) .
- 2 – The on site wind tunnel is calibrated with a pitot tube, which in turn was calibrated with a NIST traceable pitot tube.
- 3 – Ambient conditions measured with a Vaisala PTU300. Temperature, humidity and pressure readings compared to Vaisala PTU300. The temperature, pressure and humidity readings from the Vaisala® instruments were calibrated by Vaisala® against instruments calibrated to NIST traceable instruments.
- 4 – Units are calibrated at 50 knots with a 75 point correction curve.
- 5 – Unit temperature and humidity recorded with 4+ knots of wind present.
- 6 – Test conducted at a rate of turn of 70 degrees per second.

## WeatherStation® Instrument Calibration Certificate

**Model:** PB200  
**ID:** 2265415  
**Calibration Date:** 6/2/2011  
**Manufacturer :** Airmar Technology Corporation, Milford NH USA  
**Test Procedure:** 96-154-02

The WeatherStation® Instrument passes testing for IPX6 heavy seas water standards (excepting units equipped with RH sensors, which meet IPX4) and the IEC 60945 standards for exposed units<sup>1</sup>. All units are calibrated in an onsite wind tunnel<sup>2</sup>.

### Unit Test Results:

Test Performed	Logged Value	Test Requirement
0 Knot Wind	0.6	Less than 1 Knot Max
50 Knot Cal <sub>4</sub>	PASS	Within 2 Knots RMS
Humidity	N/A	Within ±5 % RH
Temperature <sub>5</sub>	22.5	Within ±1.1 °C
Pressure	997.5	Within ±2 mBar
Rate Gyro – Dynamic <sub>6</sub>	N/A	Under 4 deg/sec average error
Rate Gyro – Static	N/A	Within ±40 Deg/min
Pitch	0.0	Within 1°
Roll	0.0	Within 1°

### Ambient (Local) Test Conditions<sup>3</sup>:

**Humidity:** 39.1 %RH  
**Temperature:** 23.0 Deg C  
**Pressure:** 997.5 mBars

### Equipment List

- **Wind Tunnel Calibration**
  - United Sensor Corp. Pitot Tube – Calibrated to NIST traceable pitot tube
  - MKS Instruments Pressure Sensor
  - BK Precision® 5491A Multimeter
- **Ambient Temperature Readings (Temperature, humidity and pressure)**
  - Vaisala® PTU200
- **Rate Gyro Testing**
  - SEI A2 Absolute encoder
  - US Digital R164 Motor Controller

- 1 – After multiple heat cycles above 65° C (149°F) wind anemometer performance may require recalibration to remain within specifications at wind speeds above 50 knots (92.6 km/h) .
- 2 – The on site wind tunnel is calibrated with a pitot tube, which in turn was calibrated with a NIST traceable pitot tube.
- 3 – Ambient conditions measured with a Vaisala PTU300. Temperature, humidity and pressure readings compared to Vaisala PTU300. The temperature, pressure and humidity readings from the Vaisala® instruments were calibrated by Vaisala® against instruments calibrated to NIST traceable instruments.
- 4 – Units are calibrated at 50 knots with a 75 point correction curve.
- 5 – Unit temperature and humidity recorded with 4+ knots of wind present.
- 6 – Test conducted at a rate of turn of 70 degrees per second.