

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2251
CALIBRATION DATE: 08-Dec-15

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.03665961e+001
h = 1.36548063e+000
i = -2.24537294e-003
j = 2.21885549e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.75990	0.00000	0.00000
-1.0001	34.6768	2.79438	5.30784	2.79438	0.00000
0.9999	34.6778	2.96524	5.42490	2.96525	0.00000
14.9999	34.6789	4.25653	6.23849	4.25651	-0.00002
18.4999	34.6784	4.60204	6.43866	4.60205	0.00001
28.9999	34.6757	5.68187	7.02716	5.68187	0.00000
32.5000	34.6651	6.05260	7.21801	6.05260	-0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

